



Savyon Diagnostics Ltd.
3 Habosem Street, Ashdod
77610, Israel
Telephone: +972-8-856-2920
Fax: +972-8-852-3176

MATERIAL SAFETY DATA SHEET

QuickStripe Strep A

Catalog No. 41202

I. PRODUCT IDENTIFICATION

- | | |
|---------------|---|
| 1. Trade Name | QuickStripe Strep A |
| 2. Synonym | Qualitative test kit for the detection of Group A streptococcal antigen from throat swabs |

II. HAZARDOUS INGREDIENTS

None

III. PHYSICAL DATA

- | | |
|---------------------|---|
| 1. Appearance | Nitrocellulose strip containing anti-Strep A colloidal gold conjugate and immobilized anti-Strep A antibody |
| 2. Odor | N/A |
| 3. Boiling Point | N/A |
| 4. Evaporation Rate | N/A |
| 5. Specific Gravity | N/A |
| 6. Percent Volatile | N/A |
| 7. Vapor Density | N/A |
| 8. Vapor Pressure | N/A |
| 9. Water Solubility | N/A |

IV. FIRE AND EXPLOSION HAZARD DATA

- | | |
|-------------------------------------|------|
| 1. Flash Point | N/A |
| 2. Extinguishing Media | None |
| 3. Special Fire Fighting Procedures | None |
| 4. Unusual Fire or Explosion Hazard | None |



V. HEALTH HAZARDS

- | | |
|-----------------------------------|--|
| 1. Threshold Limits | None Suggested |
| 2. Routes of Exposure | |
| a. Ingestion | See Section 4 |
| b. Inhalation | See Section 4 |
| c. Skin Contact | See Section 4 |
| d. Skin Absorption | See Section 4 |
| e. Eye Contact | See Section 4 |
| 3. Effects of Overexposure | |
| a. Acute Overexposure | None |
| b. Chronic Overexposure | None |
| 4. Emergency First Aid Procedures | |
| a. Ingestion | N/A |
| b. Eyes | Flush affected area with copious amounts of water. Contact physician for prophylactic treatment. (This would apply only after the device has been used.) |
| c. Skin Contact | Wash affected areas thoroughly with an established germicidal agent. Contact a physician. (This would apply only after device has been used.) |
| d. Inhalation | None. |

the

VI. REACTIVITY

- | | |
|-------------------------------------|---------------------------------|
| 1. Stability conditions to avoid | (Stable <u>X</u> , Unstable _) |
| 2. Incompatibility | None |
| 3. Hazardous decomposition products | None |

VII. SPILL AND DISPOSAL PROCEDURES

- | | |
|-------------|--|
| 1. Spills | N/A to device, would only apply to samples used - biohazard precautions would apply. Treat cleaned area with established germicidal agent. |
| 2. Disposal | After use, autoclave or incinerate per local, state or federal regulations. Treat as biohazardous material |



VIII. PROTECTIVE INFORMATION

- | | |
|-------------------------|---|
| 1. Respiratory | None |
| 2. Ventilation | None |
| 3. Protective Equipment | None. Appropriate protective equipment such as gloves, face protection and lab coats should be worn when handling specimens to be tested. |

IX. SPECIAL PRECAUTIONS (HANDLING AND STORAGE)

Store at room temperature, do not freeze.

X. MISCELLANEOUS INFORMATION

Handle used device with care and consider potentially capable of transmitting infectious disease.

For further information contact:

Savyon Diagnostics Ltd.
3 Habosem Street
Ashdod, 77610
Israel
Tel: +972-8-856-2920
Fax: +972-8-852-3176